
Subject: Re: HV cable problem

Posted by [Clemens Adler](#) on Fri, 14 Jan 2005 09:01:12 GMT

[View Forum Message](#) <> [Reply to Message](#)

Hello Chamber builders,

As reported at the last status meeting I got some 300m of HV cable from CERN in December, which behaved nicely: only at 6kV the cable started to draw some 10 nA. This cable was produced in week 42 of 2002, so it was rather old. (the cable we used so far was produced in 2003).

Now I got 700m more, unfortunately not the same production date as the good one from December but produced in 2004. I tested it and it seems to be OK up to 3kV at least, above it starts to draw some currents.

Again I would like to emphasize that it seems that those currents are not spraks trough the isolator but rather surface currents over the isolator at the ends of the cables. Therefore great care should be taken that the surface of the isolator is keep clean, possibly untouched during prepareing the cables. The manufacturer recommended clean the isolation with Alcohol, or benzin, but this never helped in our case.

cheers,
Clemens
