Subject: Re: Drift HV to Radiator connection Posted by Chilo Garabatos on Mon, 14 Feb 2005 17:52:04 GMT

View Forum Message <> Reply to Message

Dear Clemens and all,

How about modifying the filter for the drift HV so that when one puts 2 kV there are 2 kV at the electrode javascript: insertTag(document.post_form.msg_body, ", ' '); ?

I spoke to Mircea and he agrees in putting a resistor there which is small enough to make a negligible deltaV. Not to have it like this is unusual and weird, I think.

Regards,

Chilo