

---

Subject: Re: FRS Calibration Issue

Posted by [mlcortes](#) on Wed, 17 Sep 2014 12:34:52 GMT

[View Forum Message](#) <> [Reply to Message](#)

---

Hi Scott!

Is not naive, they should be in the right place. Looking fast at it I see 2 possible things:

First, the  $Z_{in}$  should be the  $Z$  of the element you used to make the calibration, so if you used Ni beam  $Z_{in}$  is 28, no matter what isotope you are looking at.

Second, did you put the correct values of the magnetic fields in the  $B_{fields}$  file? Each setting has different B fields, that should be written in the runsheets. Every time you change to a different isotope, you have to change this values.

Also, just to double check that your calibration is good, run again the files you used to calibrate (putting the B fields used in each calibration file) and check that you get Ni in the right place.

Let me know if it works and if you still have the double blobs...

Liliana

---