

---

Subject: Re: FRS Calibration Issue

Posted by [mlcortes](#) on Tue, 17 Feb 2015 14:31:41 GMT

[View Forum Message](#) <> [Reply to Message](#)

---

Hallo again,

Checking the crates, I think is very likely that the problem was to reject positive values. I just removed the condition and pushed the file. You can just pull, compile and run again. After that, let me see your diff\_visu for both right and left PMTS from your production run. If this works fine, your calibration will change and it means that something in the MhTDCs was changed between the calibration and the run. One should check when the change was and how it affects you.

Let me know if it goes fine.

Liliana

---