
Subject: Re: Time protocols used for SADC

Posted by [Igor Konorov](#) on Thu, 06 May 2004 14:03:06 GMT

[View Forum Message](#) <> [Reply to Message](#)

the theses can be found in

http://www.e18.physik.tu-muenchen.de/research/publications/compass/Data_Acquisition_and_Fast_Signal_Processing_for_a_new_Positron_Emission_Tomograph.pdf

1. the TCS controller sends RESET every 14 seconds. This time interval came from SPS duty cycle.
2. there are different implementations of the counter. In current version it has 30 bits.