
Subject: Re: Forward calorimeter

Posted by [Aleksandra Wronska](#) on Tue, 19 Jun 2007 08:56:46 GMT

[View Forum Message](#) <> [Reply to Message](#)

Hi Stefano,

I fitted the calibration points from TPR with a combination of 2 exponents and got the following parameter values:

```
TF1* fun = new TF1("fun","[0]*exp([1]*x)+[2]*exp([3]*x)",0.01,3.0);  
fun->SetParameter(0,3.45);  
fun->SetParameter(1,-4.82e-03);  
fun->SetParameter(2,2.52e-01 );  
fun->SetParameter(3,-5.1914);
```

The function is different that that in TPR, but is not worse to describe the data.

cheers,
ola
