Subject: list of variables going into the offline conditions database Posted by Dariusz Miskowiec on Mon, 11 Dec 2006 15:01:50 GMT

View Forum Message <> Reply to Message

Hello,

This is the current list of variables that should go to the offline conditions database. The list still needs to be discussed (and, in my opinion, reduced).

Dariusz

File Attachments

1) nic, downloaded 996 times

Subject: Re: list of variables going into the offline conditions database Posted by Ken Oyama on Fri, 15 Dec 2006 09:12:29 GMT

View Forum Message <> Reply to Message

Hello Dariusz,

If you want to reduce the number of items:

- 1. Clock frequency is fixed to be 10.02 MHz and it will never change.
- 2. status byte of pad is not necessary to be such big. I think pad status will not have more than 10 states. 4 bits (16 states) are fine. I would put the pad status information into MCM status by extending number of MCM status bits to be 64 bits for example. It is better also because "pad status" normally related to ADC status. MCM has 21 ADC channels but if you say "pad", it can contain only 18 channels. How about to use 21 \* 4 bits = 48 bits outof 64 bits of MCM status bit?
- 3. FEE gain also required to have ADC granurality, means 21 channels per MCM. Therefore it can also be included into MCM status by extending MCM status more. It can not be chamber-by-chamber because we will probably apply lower gain for higher noise pads.
- 4. What is ADC thresholds? Perhaps it should be more precisely defined?

best regards, Ken

Subject: updated list of variables going into the offline condition database Posted by Dariusz Miskowiec on Wed, 24 Jan 2007 09:33:10 GMT View Forum Message <> Reply to Message

This is the updated list of variables which go into the offline condition database, resulting from yesterdays discussion between Raphaelle, Tom, Christoph, Christian, Anton, and Dariusz. The total parameter volume for one run is about 10 MB.

We decided not to switch from pads (18 per MCM) to ADC channels (21 per MCM) (sorry Ken).

This would require too many changes in the software. The additional ADC channels, not entering the digits, can be inspected by analyzing raw data. Pad status can be at least dead, noisy, bridged to the left, bridged to the right.

We removed the supermodule and MCM status bytes. Dead MCM's will be marked by setting the dead bit of all the involved pads.

Config summary file will be generated while configuring the electronics, and put on the file exchange server FXE. The shuttle should get it from there and translate to the individual variables needed for the offline CDB. For the unlike case that more information about the configuration is needed the variable "config id" is a name of the complete configuration file which will be stored in a subversion repository. Tom will define the format of the config summary file, provide the reading and writing software, and provide an example.

## Dariusz

## File Attachments

1) nic.txt, downloaded 1001 times

Subject: updated list of variables going into the offline conditions database Posted by Dariusz Miskowiec on Mon, 30 Jul 2007 08:11:31 GMT

View Forum Message <> Reply to Message

## File Attachments

- 1) all.txt, downloaded 1579 times
- 2) dcs.txt, downloaded 1380 times